
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 005962
7:49 13/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:52
PRESSURIZING 2:56
STERILIZING 4:00
136.5 C 221kPa 4:00
135.8 C 220kPa 4:30
136.4 C 219kPa 5:00
136.4 C 223kPa 5:30
136.4 C 222kPa 6:00
136.4 C 222kPa 6:30
136.5 C 222kPa 7:00
136.4 C 223kPa 7:30
136.4 C 223kPa 7:30
VENTING 7:30
AIR DRYING 8:18
CYCLE COMPLETE 19:11

Digital Signature #
7CB2BBBF06FB86C2
