
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.1uS / 0.7ppm
CYCLE NUMBER 001588
13:04 11/17/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:30
PRESSURIZING 7:20
STERILIZING 9:15
136.0 C 220kPa 9:15
136.0 C 219kPa 9:45
136.5 C 223kPa 10:15
136.5 C 219kPa 10:45
136.4 C 222kPa 11:15
136.4 C 222kPa 11:45
136.4 C 223kPa 12:16
136.4 C 222kPa 12:45
VENTING 12:45
AIR DRYING 13:49
CYCLE COMPLETE 17:18

Digital Signature #
63BB1FD6EBD22884
