
Midmark M11 - v1.0.5
Total Cycles: 3992

Name: RITTER2
Sterilizer ID: V2313228
10/03/2024 09:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.1 F 0.0
2:00 70.3 F 0.0
4:00 74.2 F 0.0
6:00 110.0 F 0.1
8:00 139.0 F 0.2
10:00 169.2 F 0.4
12:00 193.5 F 2.8
14:00 211.2 F 6.1
16:00 219.0 F 9.5
18:00 228.8 F 13.2
20:00 232.1 F 8.8
22:00 242.6 F 13.4
24:00 240.2 F 11.0
26:00 236.7 F 9.2
28:00 250.0 F 15.4
30:00 260.6 F 21.4
32:00 270.4 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.0
1:00 272.6 F 29.0
2:00 272.7 F 29.4
3:00 272.2 F 29.6
4:00 271.9 F 28.9

Min 271.7 F 28.9
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.1 F 0.0
5:00 196.4 F 0.0
10:00 179.0 F 0.0
15:00 177.5 F 0.0
20:00 178.7 F 0.0
25:00 179.1 F 0.0
30:00 182.7 F 0.0

Drying Complete

FILLING: 0:52
HEATING: 33:33
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:11:49
CYCLE COMPLETE

10/03/2024 10:14 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____