

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001317  
10:01 26/10/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:12  
PRESSURIZING 4:15  
STERILIZING 5:48  
136.0 C 221kPa 5:48  
136.4 C 224kPa 6:18  
136.3 C 223kPa 6:48  
136.4 C 223kPa 7:18  
136.3 C 223kPa 7:48  
136.4 C 223kPa 8:18  
136.2 C 222kPa 8:48  
136.2 C 222kPa 9:18  
136.2 C 222kPa 9:18  
VENTING 9:18  
AIR DRYING 10:16  
CYCLE COMPLETE 25:16  
-----

Digital Signature #  
3D39FE2C56B9AF25  
-----