
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001167
11:25 13/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:22
PRESSURIZING 3:24
STERILIZING 4:31
136.3 C 222kPa 4:31
136.3 C 225kPa 5:01
136.1 C 223kPa 5:31
136.4 C 226kPa 6:01
136.3 C 223kPa 6:31
136.4 C 226kPa 7:01
136.4 C 226kPa 7:31
136.4 C 226kPa 8:01
136.4 C 226kPa 8:01
VENTING 8:01
AIR DRYING 8:55
CYCLE COMPLETE 22:49

Digital Signature #
39F3904366A1D3DB
