
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000527
13:54 04/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:26
PRESSURIZING 3:31
STERILIZING 4:43
136.0 C 228kPa 4:43
136.4 C 233kPa 5:13
136.0 C 230kPa 5:43
136.4 C 232kPa 6:13
136.2 C 230kPa 6:43
136.3 C 231kPa 7:13
136.1 C 231kPa 7:43
136.3 C 232kPa 8:13
136.3 C 232kPa 8:13
VENTING 8:13
AIR DRYING 9:07
CYCLE COMPLETE 12:14

Digital Signature #
A3DE649655753ACA
