
Midmark M11 - v1.0.0
Total Cycles: 3287

Name: RITTER 1
Sterilizer ID: V2178198
05/25/2023 02:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.3 F 0.0
2:00 181.3 F 0.0
4:00 192.9 F 4.7
6:00 215.6 F 10.0
8:00 224.6 F 6.5
10:00 235.1 F 11.2
12:00 235.7 F 8.8
14:00 240.2 F 9.7
16:00 248.6 F 14.4
18:00 261.0 F 21.3
20:00 272.3 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.4 F 28.9
2:00 271.5 F 28.1
3:00 272.1 F 28.8
4:00 272.3 F 28.7

Min 271.2 F 28.0
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 234.0 F 0.0
10:00 234.1 F 0.0
15:00 224.1 F 0.0
20:00 223.5 F 0.0
25:00 226.1 F 0.0
30:00 225.9 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 20:30
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:02
CYCLE COMPLETE

05/25/2023 03:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____