
Midmark M11 - v1.0.0
Total Cycles: 5163

Name: RITTER 1
Sterilizer ID: V2178198
12/05/2024 11:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 145.5 F 0.0
2:00 151.8 F 0.2
4:00 179.8 F 0.7
6:00 203.1 F 1.5
8:00 216.2 F 6.2
10:00 227.6 F 10.8
12:00 230.5 F 6.9
14:00 242.0 F 12.7
16:00 240.5 F 11.0
18:00 238.1 F 9.8
20:00 252.3 F 16.5
22:00 263.8 F 23.2
24:00 271.5 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.7 F 29.2
2:00 272.1 F 28.6
3:00 271.6 F 28.4
4:00 272.7 F 29.2

Min 271.5 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.7 F 0.0
5:00 228.2 F 0.0
10:00 200.9 F 0.0
15:00 201.3 F 0.0
20:00 200.4 F 0.0
25:00 207.6 F 0.0
30:00 216.3 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 24:04
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:02:34
CYCLE COMPLETE

12/05/2024 12:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____