
Midmark M11 - v1.0.5
Total Cycles: 3796

Name: RITTER2
Sterilizer ID: V2313228
07/18/2024 08:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.6 F 0.0
2:00 72.1 F 0.0
4:00 100.9 F 0.0
6:00 111.7 F 0.1
8:00 162.8 F 0.5
10:00 180.5 F 4.0
12:00 198.3 F 7.6
14:00 216.6 F 11.3
16:00 222.7 F 5.1
18:00 230.9 F 9.3
20:00 240.2 F 12.0
22:00 238.6 F 11.0
24:00 236.5 F 9.0
26:00 249.6 F 15.5
28:00 260.8 F 21.8
30:00 271.1 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.6 F 29.5
2:00 271.5 F 29.1
3:00 272.6 F 29.5
4:00 271.4 F 28.9

Min 271.3 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 199.6 F 0.0
10:00 190.9 F 0.0
15:00 191.8 F 0.0
20:00 200.5 F 0.0
25:00 202.7 F 0.0
30:00 204.9 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 30:43
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:08:55
CYCLE COMPLETE

07/18/2024 09:40 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____