
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006574
13:23 07/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:16
PRESSURIZING 3:17
STERILIZING 4:20
136.4 C 222kPa 4:20
136.4 C 223kPa 4:50
136.4 C 222kPa 5:20
136.4 C 225kPa 5:50
136.3 C 223kPa 6:20
136.2 C 223kPa 6:50
136.2 C 220kPa 7:20
136.3 C 223kPa 7:50
136.3 C 223kPa 7:50
VENTING 7:51
AIR DRYING 8:42
CYCLE COMPLETE 9:35

Digital Signature #
E32E0B0FBA940743
