
Midmark M11 - v1.0.0
Total Cycles: 968

Name: RITTER 1
Sterilizer ID: V2178198
07/06/2021 11:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.8 F 0.0
2:00 197.2 F 3.6
4:00 214.7 F 8.2
6:00 220.4 F 10.3
8:00 232.2 F 9.0
10:00 241.0 F 10.2
12:00 243.4 F 12.5
14:00 242.0 F 11.3
16:00 254.9 F 18.0
18:00 266.5 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.4 F 28.9
2:00 271.5 F 28.5
3:00 272.5 F 28.9
4:00 271.5 F 28.3

Min 271.4 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.5 F 0.0
5:00 226.4 F 0.0
10:00 236.2 F 0.0
15:00 227.1 F 0.0
20:00 231.4 F 0.0
25:00 231.7 F 0.0
30:00 228.9 F 0.0
Drying Complete

FILLING: 1:57
HEATING: 19:33
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:45
CYCLE COMPLETE

07/06/2021 12:17 PM
ACCEPTED: X
REJECTED:
Operator: U01