
Midmark M11 - v1.0.0
Total Cycles: 3702

Name: RITTER 1
Sterilizer ID: V2178198
08/30/2023 08:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.4 F 0.0
2:00 83.9 F 0.0
4:00 83.4 F 0.0
6:00 107.5 F 1.6
8:00 137.1 F 3.8
10:00 170.9 F 6.7
12:00 182.4 F 10.0
14:00 205.0 F 10.8
16:00 214.2 F 4.9
18:00 223.4 F 8.2
20:00 232.2 F 11.9
22:00 231.7 F 7.4
24:00 242.6 F 12.4
26:00 239.5 F 10.1
28:00 251.4 F 15.9
30:00 261.7 F 21.9
32:00 271.4 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.6 F 29.0
2:00 271.3 F 28.2
3:00 272.4 F 29.1
4:00 271.8 F 28.3

Min 271.2 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.2 F 0.0
5:00 223.5 F 0.0
10:00 205.2 F 0.0
15:00 197.7 F 0.0
20:00 215.5 F 0.0
25:00 228.3 F 0.0
30:00 231.6 F 0.0

Drying Complete

FILLING: 0:46
HEATING: 32:42
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:10:54
CYCLE COMPLETE

08/30/2023 09:23 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____