
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 007217
13:43 17/04/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:33
PRESSURIZING 2:37
STERILIZING 3:36
136.4 C 220kPa 3:36
136.2 C 223kPa 4:06
136.4 C 223kPa 4:36
136.4 C 222kPa 5:06
136.3 C 223kPa 5:36
135.9 C 223kPa 6:06
136.3 C 223kPa 6:36
136.4 C 223kPa 7:06
136.4 C 223kPa 7:06
VENTING 7:07
AIR DRYING 8:01
CYCLE COMPLETE 8:23

Digital Signature #
82F1B66827533A2A
