

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001937  
8:31 06/07/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:50  
PRESSURIZING 3:50  
STERILIZING 4:57  
136.0 C 220kPa 4:57  
136.2 C 222kPa 5:27  
136.0 C 221kPa 5:57  
136.2 C 220kPa 6:27  
136.4 C 222kPa 6:57  
136.4 C 223kPa 7:27  
136.4 C 221kPa 7:57  
136.4 C 222kPa 8:27  
136.4 C 222kPa 8:27  
VENTING 8:27  
AIR DRYING 9:21  
CYCLE COMPLETE 9:23

-----  
Digital Signature #  
541ED6FA463B8478  
-----