
Midmark M11 - v1.0.0
Total Cycles: 4475

Name: RITTER 1
Sterilizer ID: V2178198
04/22/2024 08:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.5 F 0.0
2:00 93.9 F 0.1
4:00 88.7 F 0.5
6:00 111.5 F 2.6
8:00 146.1 F 5.4
10:00 172.7 F 8.8
12:00 201.0 F 12.3
14:00 212.6 F 4.0
16:00 219.2 F 7.4
18:00 225.9 F 11.2
20:00 228.5 F 6.4
22:00 240.9 F 11.2
24:00 235.1 F 8.3
26:00 248.7 F 14.5
28:00 260.2 F 20.7
30:00 270.1 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.1
1:00 272.7 F 28.7
2:00 272.1 F 28.5
3:00 272.0 F 28.9
4:00 272.6 F 29.2

Min 271.0 F 28.0
Max 273.0 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.1 F 0.0
5:00 231.2 F 0.0
10:00 233.0 F 0.0
15:00 230.9 F 0.0
20:00 224.9 F 0.0
25:00 228.1 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:41
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:08:52
CYCLE COMPLETE

04/22/2024 09:39 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____