
Midmark M11 - v1.0.5
Total Cycles: 3986

Name: RITTER2
Sterilizer ID: V2313228
10/02/2024 08:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.9 F 0.0
2:00 70.1 F 0.0
4:00 79.1 F 0.0
6:00 113.7 F 0.0
8:00 144.0 F 0.3
10:00 171.6 F 0.5
12:00 190.7 F 2.8
14:00 208.1 F 6.1
16:00 218.6 F 9.5
18:00 229.6 F 13.2
20:00 229.3 F 8.1
22:00 239.5 F 12.6
24:00 237.9 F 9.9
26:00 237.6 F 9.1
28:00 247.6 F 14.3
30:00 258.7 F 20.3
32:00 268.6 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.5 F 29.2
2:00 271.6 F 29.0
3:00 272.4 F 29.1
4:00 271.6 F 29.0

Min 271.3 F 28.6
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.2 F 0.0
5:00 232.6 F 0.0
10:00 221.0 F 0.0
15:00 187.0 F 0.0
20:00 174.8 F 0.0
25:00 173.2 F 0.0
30:00 169.5 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 33:14
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:11:23
CYCLE COMPLETE

10/02/2024 09:51 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____