
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006272
8:39 11/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:45
PRESSURIZING 3:46
STERILIZING 4:56
136.5 C 220kPa 4:56
136.2 C 222kPa 5:26
136.5 C 223kPa 5:56
136.4 C 223kPa 6:26
136.2 C 223kPa 6:56
136.0 C 220kPa 7:26
136.3 C 219kPa 7:56
136.3 C 223kPa 8:26
136.3 C 223kPa 8:26
VENTING 8:26
AIR DRYING 9:20
CYCLE COMPLETE 39:20

Digital Signature #
547C6AD54C1BCA82
