
Midmark M11 - v1.0.0
Total Cycles: 4725

Name: RITTER 1
Sterilizer ID: V2178198
07/12/2024 02:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.2 F 0.0
2:00 152.1 F 3.1
4:00 174.4 F 6.7
6:00 193.5 F 10.9
8:00 221.0 F 4.0
10:00 222.8 F 9.5
12:00 231.3 F 9.2
14:00 239.3 F 11.1
16:00 236.5 F 8.7
18:00 250.8 F 15.5
20:00 262.3 F 22.3
22:00 272.3 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 271.7 F 28.6
2:00 272.2 F 28.7
3:00 272.4 F 28.9
4:00 272.3 F 29.1

Min 271.5 F 28.6
Max 273.0 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.3 F 0.0
5:00 229.0 F 0.0
10:00 207.3 F 0.0
15:00 193.5 F 0.0
20:00 182.7 F 0.0
25:00 181.8 F 0.0
30:00 181.2 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 22:48
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:01:20
CYCLE COMPLETE

07/12/2024 03:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____