
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006324
13:43 18/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:44
PRESSURIZING 1:50
STERILIZING 2:51
136.5 C 220kPa 2:51
136.0 C 223kPa 3:21
136.1 C 222kPa 3:51
136.4 C 225kPa 4:21
136.1 C 218kPa 4:51
136.2 C 221kPa 5:21
136.1 C 217kPa 5:51
136.4 C 224kPa 6:21
136.4 C 224kPa 6:21
VENTING 6:22
AIR DRYING 7:12
CYCLE COMPLETE 9:25

Digital Signature #
0A5ECF39384CDE53
