

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2523

Name: RITTER 1  
Sterilizer ID: V2178198  
11/07/2022 06:02 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 170.7 F 0.0  
2:00 168.0 F 3.9  
4:00 187.0 F 8.4  
6:00 207.4 F 13.5  
8:00 223.9 F 7.4  
10:00 236.1 F 12.7  
12:00 237.4 F 10.1  
14:00 235.7 F 8.1  
16:00 250.5 F 15.5  
18:00 262.9 F 22.7  
20:00 271.8 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.6 F 29.0  
2:00 271.4 F 28.3  
3:00 272.6 F 29.2  
4:00 272.0 F 28.5

Min 271.4 F 28.2  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.4 F 0.0  
5:00 237.3 F 0.0  
10:00 233.4 F 0.0  
15:00 227.5 F 0.0  
20:00 229.1 F 0.0  
25:00 230.0 F 0.0  
30:00 228.9 F 0.0  
Drying Complete

FILLING: 1:13  
HEATING: 20:07  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:58:28  
CYCLE COMPLETE

11/07/2022 07:00 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_