
Midmark M11 - v1.0.5
Total Cycles: 2200

Name: RITTER2
Sterilizer ID: V2313228
03/30/2023 01:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.1 F 0.0
2:00 182.1 F 4.1
4:00 193.4 F 8.9
6:00 213.0 F 7.8
8:00 225.4 F 9.7
10:00 231.8 F 8.3
12:00 242.6 F 12.6
14:00 244.9 F 13.0
16:00 257.9 F 20.1
18:00 269.4 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.3 F 29.6
2:00 271.2 F 29.0
3:00 272.3 F 29.8
4:00 271.8 F 29.2

Min 271.2 F 28.9
Max 272.4 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.5 F 0.0
5:00 221.0 F 0.0
10:00 240.2 F 0.0
15:00 230.9 F 0.0
20:00 226.2 F 0.0
25:00 228.0 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 18:59
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:57:54
CYCLE COMPLETE

03/30/2023 02:07 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____