

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1473

Name: RITTER 1  
Sterilizer ID: V2178198  
12/20/2021 09:30 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 197.5 F 0.0  
2:00 193.0 F 2.4  
4:00 212.6 F 7.0  
6:00 222.0 F 12.6  
8:00 223.4 F 7.8  
10:00 236.7 F 11.8  
12:00 239.9 F 11.1  
14:00 238.1 F 9.3  
16:00 253.0 F 16.5  
18:00 265.3 F 23.7  
20:00 272.4 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.7  
1:00 272.4 F 28.9  
2:00 272.8 F 29.4  
3:00 272.1 F 28.6  
4:00 272.9 F 29.2

Min 271.8 F 28.6  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.8 F 0.0  
5:00 219.4 F 0.0  
10:00 228.0 F 0.0  
15:00 223.9 F 0.0  
20:00 229.8 F 0.0  
25:00 229.9 F 0.0  
30:00 230.2 F 0.0  
Drying Complete

FILLING: 1:21  
HEATING: 20:28  
STERILIZING: 4:00  
VENTING: 2:54  
DRYING: 30:00  
TOTAL CYCLE: 00:58:43  
CYCLE COMPLETE

12/20/2021 10:30 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02