
Midmark M11 - v1.0.0
Total Cycles: 1214

Name: RITTER 1
Sterilizer ID: V2178198
09/28/2021 02:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.9 F 0.0
2:00 179.4 F 2.9
4:00 194.1 F 7.5
6:00 216.5 F 13.1
8:00 225.8 F 8.1
10:00 238.6 F 13.1
12:00 239.7 F 11.2
14:00 238.1 F 9.6
16:00 252.4 F 16.6
18:00 264.4 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.7 F 29.0
2:00 271.7 F 28.4
3:00 272.8 F 29.2
4:00 272.0 F 28.5

Min 271.6 F 28.3
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.4 F 0.0
5:00 215.7 F 0.0
10:00 225.3 F 0.0
15:00 233.6 F 0.0
20:00 218.0 F 0.0
25:00 208.5 F 0.0
30:00 202.7 F 0.0
Drying Complete

FILLING: 1:52
HEATING: 19:54
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:58:50
CYCLE COMPLETE

09/28/2021 03:41 PM
ACCEPTED: X
REJECTED:
Operator: U01