

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1206

Name: RITTER2  
Sterilizer ID: V2313228  
06/21/2022 10:33 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.8 F 0.0  
2:00 172.7 F 3.2  
4:00 187.2 F 8.0  
6:00 222.5 F 12.5  
8:00 226.3 F 9.0  
10:00 234.6 F 10.0  
12:00 242.5 F 12.5  
14:00 241.7 F 11.6  
16:00 255.0 F 18.4  
18:00 266.5 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 272.4 F 29.3  
1:00 272.4 F 29.5  
2:00 272.3 F 29.8  
3:00 271.6 F 29.3  
4:00 272.1 F 29.5  
  
Min 271.5 F 29.1  
Max 272.6 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.8 F 0.0  
5:00 227.7 F 0.0  
10:00 226.6 F 0.0  
15:00 183.5 F 0.0  
20:00 181.1 F 0.0  
25:00 182.6 F 0.0  
30:00 178.0 F 0.0  
Drying Complete

FILLING: 2:11  
HEATING: 19:21  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:58:44  
CYCLE COMPLETE

06/21/2022 11:31 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_