

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4132

Name: RITTER 1  
Sterilizer ID: V2178198  
12/18/2023 03:14 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 128.1 F 0.0  
2:00 129.0 F 1.7  
4:00 149.1 F 4.6  
6:00 175.2 F 8.4  
8:00 201.0 F 13.4  
10:00 218.7 F 5.9  
12:00 228.2 F 10.0  
14:00 233.1 F 8.9  
16:00 239.2 F 11.1  
18:00 235.3 F 8.8  
20:00 248.9 F 15.0  
22:00 260.2 F 21.2  
24:00 270.5 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.6  
1:00 272.4 F 29.4  
2:00 271.6 F 28.6  
3:00 272.6 F 29.4  
4:00 271.8 F 28.7

Min 271.2 F 28.5  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.6 F 0.0  
5:00 232.1 F 0.0  
10:00 230.5 F 0.0  
15:00 225.8 F 0.0  
20:00 224.6 F 0.0  
25:00 233.7 F 0.0  
30:00 225.8 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 24:43  
STERILIZING: 4:00  
VENTING: 3:29  
DRYING: 30:00  
TOTAL CYCLE: 01:03:18  
CYCLE COMPLETE

12/18/2023 04:17 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_