
Midmark M11 - v1.0.5
Total Cycles: 1175

Name: RITTER2
Sterilizer ID: V2313228
06/08/2022 12:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.1 F 0.0
2:00 183.8 F 3.7
4:00 189.5 F 8.3
6:00 209.2 F 11.9
8:00 224.2 F 8.8
10:00 235.1 F 11.0
12:00 240.4 F 12.2
14:00 240.2 F 11.2
16:00 253.2 F 17.8
18:00 264.7 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.6 F 30.0
2:00 271.8 F 29.3
3:00 271.8 F 29.8
4:00 272.5 F 29.9

Min 271.4 F 29.1
Max 272.6 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 215.7 F 0.0
10:00 245.3 F 0.0
15:00 240.6 F 0.0
20:00 233.4 F 0.0
25:00 217.9 F 0.0
30:00 215.0 F 0.0

Drying Complete

FILLING: 1:27
HEATING: 19:55
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:33
CYCLE COMPLETE

06/08/2022 01:53 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____