

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1195

Name: RITTER2  
Sterilizer ID: V2313228  
06/16/2022 10:26 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 149.2 F 0.0  
2:00 149.5 F 2.4  
4:00 166.9 F 6.2  
6:00 190.1 F 10.9  
8:00 213.9 F 4.8  
10:00 226.6 F 10.0  
12:00 230.8 F 8.6  
14:00 240.9 F 12.5  
16:00 239.5 F 11.1  
18:00 252.5 F 17.5  
20:00 263.4 F 24.0  
22:00 272.0 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.3  
1:00 271.6 F 29.8  
2:00 272.1 F 29.9  
3:00 271.3 F 29.2  
4:00 271.5 F 29.7

Min 271.0 F 29.1  
Max 272.7 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.4 F 0.0  
5:00 239.7 F 0.0  
10:00 227.8 F 0.0  
15:00 221.6 F 0.0  
20:00 215.3 F 0.0  
25:00 207.7 F 0.0  
30:00 185.8 F 0.0  
Drying Complete

FILLING: 1:35  
HEATING: 22:02  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:00:46  
CYCLE COMPLETE

06/16/2022 11:26 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_