
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
2.6uS / 1.6ppm
CYCLE NUMBER 003698
13:40 26/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:19
PRESSURIZING 2:31
STERILIZING 3:23
136.0 C 220kPa 3:23
135.7 C 220kPa 3:53
136.2 C 224kPa 4:23
136.2 C 222kPa 4:53
136.1 C 222kPa 5:23
136.4 C 225kPa 5:53
136.4 C 220kPa 6:23
136.3 C 223kPa 6:53
136.3 C 223kPa 6:53
VENTING 6:53
AIR DRYING 7:43
CYCLE COMPLETE 8:25

Digital Signature #
79ED4223F22EC853
