
Midmark M11 - v1.0.5
Total Cycles: 575

Name: RITTER2
Sterilizer ID: V2313228
11/24/2021 03:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.3 F 0.0
2:00 133.2 F 1.5
4:00 137.6 F 4.6
6:00 168.6 F 8.9
8:00 195.6 F 9.1
10:00 213.6 F 7.4
12:00 227.9 F 12.4
14:00 232.8 F 9.2
16:00 240.2 F 10.6
18:00 244.3 F 13.3
20:00 255.9 F 19.5
22:00 266.6 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.5
1:00 272.7 F 30.3
2:00 272.2 F 29.7
3:00 271.7 F 29.7
4:00 272.8 F 30.3

Min 271.5 F 29.4
Max 272.8 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.5 F 0.0
5:00 194.5 F 0.0
10:00 201.2 F 0.0
15:00 200.2 F 0.0
20:00 205.3 F 0.0
25:00 203.7 F 0.0
30:00 206.1 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 23:40
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:02:12
CYCLE COMPLETE

11/24/2021 04:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____