
Midmark M11 - v1.0.0
Total Cycles: 2098

Name: RITTER 1
Sterilizer ID: V2178198
07/07/2022 10:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.1 F 0.0
2:00 178.0 F 2.3
4:00 184.3 F 6.3
6:00 216.3 F 11.5
8:00 225.0 F 5.9
10:00 235.4 F 11.0
12:00 235.5 F 7.7
14:00 243.8 F 12.1
16:00 244.7 F 12.7
18:00 257.8 F 19.6
20:00 269.2 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.6 F 29.0
2:00 271.6 F 28.4
3:00 272.6 F 29.1
4:00 271.8 F 28.3

Min 271.5 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.9 F 0.0
5:00 213.2 F 0.0
10:00 226.2 F 0.0
15:00 231.1 F 0.0
20:00 224.6 F 0.0
25:00 224.5 F 0.0
30:00 220.6 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 21:03
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:58
CYCLE COMPLETE

07/07/2022 11:41 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____