

-----  
Midmark M11 - v1.0.5  
Total Cycles: 300

Name: RITTER2  
Sterilizer ID: V2313228  
08/23/2021 09:12 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.8 F 0.0  
2:00 179.1 F 3.4  
4:00 192.9 F 8.2  
6:00 210.5 F 10.6  
8:00 226.4 F 8.8  
10:00 236.8 F 11.0  
12:00 242.4 F 12.6  
14:00 242.6 F 12.0  
16:00 255.9 F 19.0  
18:00 267.5 F 26.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.9  
1:00 272.3 F 29.3  
2:00 272.1 F 29.6  
3:00 272.0 F 29.0  
4:00 272.2 F 29.7  
  
Min 271.4 F 28.8  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.4 F 0.0  
5:00 231.8 F 0.0  
10:00 212.9 F 0.0  
15:00 210.7 F 0.0  
20:00 205.0 F 0.0  
25:00 202.0 F 0.0  
30:00 200.1 F 0.0  
Drying Complete

FILLING: 1:57  
HEATING: 19:20  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 00:58:21  
CYCLE COMPLETE

08/23/2021 10:12 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02