
Midmark M11 - v1.0.5
Total Cycles: 2966

Name: RITTER2
Sterilizer ID: V2313228
10/04/2023 11:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.0 F 0.0
2:00 164.7 F 2.0
4:00 186.4 F 6.0
6:00 204.9 F 10.8
8:00 220.9 F 5.6
10:00 231.4 F 10.8
12:00 232.2 F 7.6
14:00 242.7 F 12.6
16:00 244.3 F 12.6
18:00 256.2 F 18.9
20:00 266.9 F 25.5
22:00 273.2 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 271.4 F 29.3
2:00 271.8 F 29.3
3:00 272.0 F 29.3
4:00 272.6 F 29.5

Min 271.3 F 29.2
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.5 F 0.0
5:00 229.8 F 0.0
10:00 227.9 F 0.0
15:00 227.6 F 0.0
20:00 229.2 F 0.0
25:00 229.1 F 0.0
30:00 183.8 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 22:14
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:01:09
CYCLE COMPLETE

10/04/2023 12:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____