
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001384
11:16 16/12/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 0:18
STOP BUTTON PRESSED
NOT STERILE
END TIME 0:18

Digital Signature #
A60A3DDE84525733
