
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001587
14:57 13/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:44
PRESSURIZING 2:50
STERILIZING 3:54
136.4 C 224kPa 3:54
136.2 C 223kPa 4:24
136.3 C 223kPa 4:54
136.2 C 224kPa 5:24
136.2 C 221kPa 5:54
136.3 C 223kPa 6:24
136.2 C 223kPa 6:54
136.4 C 225kPa 7:24
136.4 C 225kPa 7:24
VENTING 7:24
AIR DRYING 8:13
CYCLE COMPLETE 8:21

Digital Signature #
422365B5ED7F8753
