
Midmark M11 - v1.0.5
Total Cycles: 440

Name: RITTER2
Sterilizer ID: V2313228
10/12/2021 01:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.7 F 0.0
2:00 190.1 F 3.5
4:00 198.9 F 9.4
6:00 216.2 F 6.7
8:00 230.6 F 11.5
10:00 235.4 F 9.8
12:00 236.8 F 9.0
14:00 250.3 F 16.0
16:00 262.3 F 23.0
18:00 272.6 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.2
1:00 272.3 F 29.6
2:00 271.7 F 29.7
3:00 272.0 F 29.4
4:00 272.0 F 29.9

Min 271.1 F 29.0
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.2 F 0.0
5:00 228.9 F 0.0
10:00 227.4 F 0.0
15:00 214.7 F 0.0
20:00 238.7 F 0.0
25:00 232.8 F 0.0
30:00 195.8 F 0.0
Drying Complete

FILLING: 1:52
HEATING: 18:15
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:57:11
CYCLE COMPLETE

10/12/2021 02:14 PM
ACCEPTED: X
REJECTED:
Operator: U01