
Midmark M11 - v1.0.0
Total Cycles: 4149

Name: RITTER 1
Sterilizer ID: V2178198
12/21/2023 03:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 126.9 F 0.0
2:00 130.8 F 1.1
4:00 151.4 F 3.6
6:00 159.2 F 6.9
8:00 182.2 F 10.5
10:00 214.6 F 5.2
12:00 215.1 F 6.1
14:00 227.9 F 10.1
16:00 233.6 F 9.1
18:00 239.2 F 10.9
20:00 235.4 F 8.2
22:00 248.3 F 14.3
24:00 258.8 F 20.0
26:00 268.9 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.1
1:00 272.4 F 28.9
2:00 272.0 F 28.4
3:00 271.0 F 28.1
4:00 272.3 F 28.9

Min 271.0 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.4 F 0.0
5:00 202.0 F 0.0
10:00 226.1 F 0.0
15:00 225.5 F 0.0
20:00 227.2 F 0.0
25:00 230.4 F 0.0
30:00 234.4 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 27:14
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:06:06
CYCLE COMPLETE

12/21/2023 04:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____