
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 008287
12:58 31/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:57
PRESSURIZING 1:59
STERILIZING 3:04
136.4 C 221kPa 3:04
136.2 C 223kPa 3:34
136.4 C 224kPa 4:04
136.2 C 223kPa 4:34
136.4 C 224kPa 5:04
136.4 C 221kPa 5:34
136.4 C 225kPa 6:04
136.2 C 223kPa 6:34
136.2 C 223kPa 6:34
VENTING 6:34
AIR DRYING 7:25
CYCLE COMPLETE 29:52

Digital Signature #
578B4536663DE569
