
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001429
11:28 14/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:54
PRESSURIZING 3:55
STERILIZING 4:55
136.4 C 224kPa 4:55
136.3 C 224kPa 5:25
136.2 C 221kPa 5:55
136.2 C 224kPa 6:25
136.4 C 225kPa 6:55
136.4 C 223kPa 7:25
136.3 C 224kPa 7:55
136.3 C 224kPa 8:25
136.3 C 224kPa 8:25
VENTING 8:26
AIR DRYING 9:19
CYCLE COMPLETE 13:31

Digital Signature #
14953F3F1640C586
