
Midmark M11 - v1.0.5
Total Cycles: 1310

Name: RITTER2
Sterilizer ID: V2313228
07/18/2022 09:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.2 F 0.0
2:00 150.7 F 0.0
4:00 157.5 F 0.2
6:00 196.2 F 2.4
8:00 210.1 F 6.7
10:00 223.7 F 11.4
12:00 228.0 F 6.6
14:00 238.8 F 11.7
16:00 235.8 F 9.3
18:00 238.9 F 9.7
20:00 248.1 F 14.7
22:00 259.9 F 21.4
24:00 270.7 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.3 F 29.5
2:00 271.5 F 29.2
3:00 272.2 F 29.4
4:00 271.4 F 29.1

Min 271.3 F 28.9
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.2 F 0.0
5:00 218.9 F 0.0
10:00 242.5 F 0.0
15:00 238.1 F 0.0
20:00 225.0 F 0.0
25:00 225.4 F 0.0
30:00 223.5 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 24:46
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:02:51
CYCLE COMPLETE

07/18/2022 10:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____