

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1500

Name: RITTER2  
Sterilizer ID: V2313228  
09/19/2022 10:09 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 201.9 F 0.0  
2:00 191.2 F 3.5  
4:00 199.0 F 8.7  
6:00 225.6 F 8.5  
8:00 231.1 F 10.7  
10:00 234.8 F 8.6  
12:00 239.1 F 9.9  
14:00 248.9 F 14.9  
16:00 261.4 F 22.1  
18:00 272.8 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.6  
1:00 272.1 F 29.5  
2:00 272.4 F 29.7  
3:00 272.7 F 30.0  
4:00 272.0 F 30.2  
  
Min 271.6 F 29.5  
Max 272.8 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.8 F 0.0  
5:00 228.9 F 0.0  
10:00 224.0 F 0.0  
15:00 225.5 F 0.0  
20:00 225.2 F 0.0  
25:00 227.2 F 0.0  
30:00 227.4 F 0.0  
  
Drying Complete

FILLING: 2:14  
HEATING: 19:09  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:58:32  
CYCLE COMPLETE

09/19/2022 11:07 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_