
Midmark M11 - v1.0.0
Total Cycles: 5426

Name: RITTER 1
Sterilizer ID: V2178198
03/20/2025 11:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 85.5 F 0.0
2:00 87.0 F 0.6
4:00 108.0 F 2.6
6:00 140.8 F 5.3
8:00 153.4 F 8.0
10:00 169.8 F 11.3
12:00 199.1 F 3.5
14:00 214.8 F 6.2
16:00 219.4 F 9.8
18:00 236.2 F 10.2
20:00 234.2 F 9.6
22:00 238.8 F 10.0
24:00 243.1 F 12.3
26:00 254.1 F 17.9
28:00 264.1 F 23.9
30:00 272.0 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 271.9 F 29.4
2:00 271.2 F 29.1
3:00 271.9 F 29.4
4:00 271.0 F 28.9

Min 271.0 F 28.8
Max 272.0 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.9 F 0.0
5:00 193.1 F 0.0
10:00 202.2 F 0.0
15:00 206.4 F 0.0
20:00 201.8 F 0.0
25:00 204.1 F 0.0
30:00 211.7 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 30:10
STERILIZING: 4:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:08:48
CYCLE COMPLETE

03/20/2025 12:46 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____