
Midmark M11 - v1.0.0
Total Cycles: 608

Name: MIDMARK
Sterilizer ID: V2178198
01/18/2021 09:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 118.5 F 0.0
2:00 118.0 F 0.0
4:00 151.9 F 0.7
6:00 164.8 F 4.0
8:00 187.2 F 8.5
10:00 210.6 F 13.5
12:00 225.6 F 8.7
14:00 236.6 F 10.4
16:00 240.0 F 11.8
18:00 239.3 F 10.0
20:00 253.2 F 16.8
22:00 265.1 F 23.9
24:00 273.1 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.9
1:00 272.6 F 29.0
2:00 271.5 F 29.0
3:00 272.5 F 29.0
4:00 271.8 F 28.8

Min 271.4 F 28.7
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.8 F 0.0
5:00 230.7 F 0.0
10:00 221.2 F 0.0
15:00 219.3 F 0.0
20:00 222.0 F 0.0
25:00 230.3 F 0.0
30:00 231.9 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 24:14
STERILIZING: 4:00
VENTING: 2:60
DRYING: 30:00
TOTAL CYCLE: 01:02:06
CYCLE COMPLETE

01/18/2021 10:41 AM
ACCEPTED: X
REJECTED:
Operator: U01