
Midmark M11 - v1.0.0
Total Cycles: 4720

Name: RITTER 1
Sterilizer ID: V2178198
07/11/2024 06:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.9 F 0.0
2:00 169.5 F 3.5
4:00 180.1 F 8.0
6:00 207.0 F 12.8
8:00 221.1 F 6.7
10:00 232.0 F 12.0
12:00 233.8 F 8.9
14:00 239.3 F 9.7
16:00 246.9 F 13.9
18:00 259.7 F 20.8
20:00 271.0 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.5 F 29.4
2:00 271.6 F 28.6
3:00 271.9 F 29.2
4:00 272.2 F 29.2

Min 271.1 F 28.4
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.1 F 0.0
5:00 214.0 F 0.0
10:00 222.2 F 0.0
15:00 219.3 F 0.0
20:00 238.9 F 0.0
25:00 241.9 F 0.0
30:00 225.6 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 20:47
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 00:59:31
CYCLE COMPLETE

07/11/2024 07:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____