
Midmark M11 - v1.0.0
Total Cycles: 5186

Name: RITTER 1
Sterilizer ID: V2178198
12/13/2024 11:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.0 F 0.0
2:00 157.9 F 0.0
4:00 171.3 F 0.1
6:00 190.7 F 2.2
8:00 207.9 F 6.1
10:00 221.1 F 10.2
12:00 232.2 F 8.0
14:00 234.9 F 9.6
16:00 239.5 F 9.6
18:00 244.6 F 12.4
20:00 242.1 F 10.8
22:00 254.3 F 17.1
24:00 265.0 F 23.5
26:00 271.2 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 27.8
1:00 272.4 F 28.6
2:00 272.3 F 28.1
3:00 271.2 F 27.7
4:00 272.0 F 28.5

Min 271.2 F 27.7
Max 272.5 F 28.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.4 F 0.0
5:00 224.7 F 0.0
10:00 203.2 F 0.0
15:00 198.0 F 0.0
20:00 200.2 F 0.0
25:00 194.3 F 0.0
30:00 193.2 F 0.0
Drying Complete

FILLING: 0:43
HEATING: 26:00
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:04:23
CYCLE COMPLETE

12/13/2024 12:24 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____