

-----  
Midmark M11 - v1.0.5  
Total Cycles: 115

Name: RITTER2  
Sterilizer ID: V2313228  
06/17/2021 03:45 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 118.3 F 0.0  
2:00 118.1 F 1.0  
4:00 124.6 F 2.8  
6:00 150.2 F 5.3  
8:00 172.1 F 8.4  
10:00 190.6 F 11.7  
12:00 204.8 F 3.8  
14:00 216.8 F 7.4  
16:00 225.9 F 11.1  
18:00 230.5 F 7.9  
20:00 238.6 F 12.4  
22:00 240.2 F 11.0  
24:00 250.5 F 16.2  
26:00 260.1 F 21.6  
28:00 268.9 F 27.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.2  
1:00 271.7 F 29.8  
2:00 272.7 F 30.5  
3:00 272.3 F 30.0  
4:00 271.7 F 30.0

Min 270.9 F 29.1  
Max 273.0 F 30.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.4 F 0.0  
5:00 228.1 F 0.0  
10:00 222.5 F 0.0  
15:00 219.3 F 0.0  
20:00 215.6 F 0.0  
25:00 213.3 F 0.0  
30:00 210.7 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 29:14  
STERILIZING: 4:00  
VENTING: 3:54  
DRYING: 30:00  
TOTAL CYCLE: 01:08:02  
CYCLE COMPLETE

06/17/2021 04:53 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_