
Midmark M11 - v1.0.0
Total Cycles: 5403

Name: RITTER 1
Sterilizer ID: V2178198
03/10/2025 10:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.6 F 0.0
2:00 125.9 F 1.3
4:00 136.5 F 4.2
6:00 162.4 F 7.5
8:00 187.6 F 11.1
10:00 205.3 F 3.1
12:00 214.1 F 5.6
14:00 225.5 F 9.3
16:00 236.1 F 13.5
18:00 236.4 F 9.6
20:00 240.3 F 10.4
22:00 244.7 F 12.6
24:00 255.7 F 18.3
26:00 265.7 F 24.4
28:00 271.4 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.4 F 29.4
2:00 272.4 F 29.1
3:00 271.9 F 28.7
4:00 271.4 F 28.6

Min 271.3 F 28.5
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.6 F 0.0
5:00 197.9 F 0.0
10:00 247.5 F 0.0
15:00 231.0 F 0.0
20:00 219.4 F 0.0
25:00 224.4 F 0.0
30:00 216.7 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 28:46
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:07:27
CYCLE COMPLETE

03/10/2025 11:24 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____