

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4491

Name: RITTER2  
Sterilizer ID: V2313228  
04/07/2025 09:29 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.3 F 0.0  
2:00 81.6 F 0.0  
4:00 131.2 F 0.3  
6:00 110.3 F 2.4  
8:00 183.4 F 5.6  
10:00 166.9 F 9.0  
12:00 185.8 F 12.7  
14:00 214.4 F 3.5  
16:00 222.7 F 6.7  
18:00 231.9 F 10.5  
20:00 229.6 F 8.4  
22:00 239.1 F 11.3  
24:00 236.6 F 9.2  
26:00 249.8 F 15.5  
28:00 260.8 F 21.8  
30:00 271.0 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.0  
1:00 272.3 F 28.9  
2:00 272.5 F 29.3  
3:00 272.5 F 29.6  
4:00 271.7 F 29.1

Min 271.6 F 28.8  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.7 F 0.0  
5:00 194.1 F 0.0  
10:00 194.7 F 0.0  
15:00 193.9 F 0.0  
20:00 205.3 F 0.0  
25:00 211.1 F 0.0  
30:00 213.1 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 30:39  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:09:03  
CYCLE COMPLETE

04/07/2025 10:38 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_