
Midmark M11 - v1.0.0
Total Cycles: 5451

Name: RITTER 1
Sterilizer ID: V2178198
03/26/2025 04:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.5 F 0.0
2:00 174.8 F 3.5
4:00 192.8 F 7.7
6:00 211.4 F 12.7
8:00 222.3 F 6.4
10:00 231.5 F 11.2
12:00 233.9 F 7.9
14:00 244.4 F 12.1
16:00 243.8 F 12.5
18:00 256.5 F 18.9
20:00 267.3 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.3 F 29.4
2:00 272.3 F 28.9
3:00 271.1 F 28.6
4:00 271.9 F 29.4

Min 271.1 F 28.5
Max 272.5 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.0 F 0.0
5:00 213.2 F 0.0
10:00 209.7 F 0.0
15:00 217.7 F 0.0
20:00 210.8 F 0.0
25:00 203.1 F 0.0
30:00 208.1 F 0.0
Drying Complete

FILLING: 1:38
HEATING: 21:31
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:00:59
CYCLE COMPLETE

03/26/2025 05:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____