
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002280
11:52 18/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:05
PRESSURIZING 2:08
STERILIZING 3:11
136.0 C 220kPa 3:11
136.4 C 222kPa 3:41
136.4 C 226kPa 4:11
136.3 C 225kPa 4:41
136.4 C 226kPa 5:11
136.4 C 226kPa 5:41
136.4 C 226kPa 6:11
136.3 C 223kPa 6:41
136.3 C 223kPa 6:41
VENTING 6:41
AIR DRYING 7:34
CYCLE COMPLETE 20:04

Digital Signature #
6750FB0923A4391C
