
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 002339
15:38 29/11/2024
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 1:40
PRESSURIZING 3:48
STERILIZING 4:32
136.0 C 222kPa 4:32
136.3 C 223kPa 5:02
136.4 C 224kPa 5:32
136.4 C 226kPa 6:02
136.4 C 227kPa 6:32
136.4 C 226kPa 7:02
136.4 C 225kPa 7:32
136.3 C 224kPa 8:02
136.1 C 222kPa 8:32
136.3 C 225kPa 9:02
136.4 C 226kPa 9:32
136.4 C 226kPa 10:02
136.4 C 226kPa 10:32
136.4 C 222kPa 11:02
136.4 C 226kPa 11:32
136.4 C 223kPa 12:02
135.9 C 226kPa 12:32
136.4 C 226kPa 13:02
136.4 C 226kPa 13:32
136.4 C 227kPa 14:02
136.4 C 227kPa 14:32
136.4 C 227kPa 14:32
VENTING 14:32
AIR DRYING 15:23
CYCLE COMPLETE 21:00

Digital Signature #
CA8F3F9CC0D6B793
